

Date

Reference

2023-11-15

2023/2282

Scope of accreditation

Calibration according to SS-EN ISO/IEC 17025:2018

RTI Group AB

Mölnadal

Accreditation number

2021

A001514-001

Electricity and Magnetism

<i>Technology area</i>	<i>Method</i>	<i>Parameter</i>	<i>Material</i>	<i>Measure</i>	<i>Best measuring ability (CMC) +/-</i>	<i>Technique</i>	<i>Field</i>
Current	MTB-030, utg G, 2018	DC	Current showing	±1 pA - ± 100 pA, ±100 pA - ± 10 mA	0,1 pA, 0,1%		No
	MTB-050, utg H, 2018	DC	Current showing	0,5 – 1500 mA	0,13 %		No
Voltage	MTB-010 Utg I, 2018	DC	Multimeter	10 – 150 kV	0,56 %		No

Ionizing radiation

<i>Technology area</i>	<i>Method</i>	<i>Parameter</i>	<i>Material</i>	<i>Measure</i>	<i>Best measuring ability (CMC) +/-</i>	<i>Technique</i>	<i>Field</i>
Air kerma	MTB-020, ed L, 2023		Dosimeter	18 – 150 kV	1,62%		No
Kerma-area-product	MTB-070, ed H, 2023		Dosimeter	18 – 150 kV	3,5%	Röntgen KAP	No
Kerma-length-product	MTB-060, ed H, 2023		Dosimeter	18 – 150 kV	2%		No

Photometer and radiometer

<i>Technology area</i>	<i>Method</i>	<i>Parameter</i>	<i>Material</i>	<i>Measure</i>	<i>Best measuring ability (CMC) +/-</i>	<i>Technique</i>	<i>Field</i>
Illuminance	MTB-040, ed J, 2023		Photometer	10 – 100 lux	3,3%		No
Luminance	MTB-040, ed J, 2023		Photometer	10 – 1000 cd/m ²	3,3%		No

Calibration and measurement capability, CMC, is the smallest uncertainty the calibration laboratory can provide, expressed as the expanded uncertainty having a coverage probability of approximately 95%.

Changes in the scope of accreditation are in bold.

Date

2023-11-15

Appendix 1

Reference

2023/2282